

ABAXIS VETSCAN CHEMISTRY VERSION 2013.1 AND HIGHER

TABLE OF CONTENTS

Interfacing AVImark with the Abaxis VetScan Chemistry	3
Machine Settings.....	3
AVImark Setup	3
Use.....	5

INTERFACING AVIMARK WITH THE ABAXIS VETSCAN CHEMISTRY

Machine Settings

There are some settings that need to be made on the VetScan Chemistry to insure the communication between the machine and AVImark. The settings are as follows:

- ✓ The Abaxis VetScan Chemistry machine needs to be set to a baud rate of 9600
- ✓ Make sure the machine is connected with the cables provided by Abaxis.


AVImark Setup

From the AVImark workstation you wish to use for communication:

1. Select **Instrument Management** from the **Services** menu on the CID to open the Instruments window.
2. Open a new Instrument Setup window using a **right-click | New**.
3. Select **Abaxis VetScan Chemistry** from the available drop-down list.

The Name, Computer ID, and Protocol should automatically be populated with the proper information once the instrument has been selected.


4. Place a check in the **Active** box if not already defaulted to being checked.



5. Click **Next** and choose the appropriate Com Port. The **Baud Rate** should be set to **9600**.
6. Click **Finish** to complete the instrument setup.

AVImark has tested the ability to utilize a “Serial to USB Adapter Cable” for various lab instrument interfaces with limited success. You will need to check with your lab instrument manufacturer to see if they supply this Serial to USB Adaptor Cable or if they recommend one over another (for there are various brands and types available on the market). With this in mind, AVImark does not guarantee that they will ALL work or be recognized as a communication port by your computer workstation. AVImark is limited to the technical support we can provide if you choose to use this type of device for your lab machine to computer connection.

A storage folder which will receive the results from the VetScan Chemistry is needed. This folder can be created inside the AVImark folder on the server computer. Once this storage folder is created, the path must be defined within the **Hospital Setup | Advanced Options | Save Lab Instrument Results To**.



*This folder name CANNOT contain spaces. We recommend a folder named **instrument**.*

Assign your instrument rotor(s) under the Labs tab of the appropriate AVImark treatments:

1. In the Treatment List right-click on desired treatment and select **Change**.
2. Select the **Laboratory** tab, then click on the **Labs** sub-tab
3. **Right-click | New**.
 - o Select Abaxis Instrument from **Lab Company** drop-down list.
 - o Select Abaxis VS2 from **Machine** drop-down.
 - o Select appropriate **Rotor** from **Rotor** drop-down.
4. Click **OK**.

The Instrument Import utility is much the same as in previous AVImark versions.

- ✓ The utility can be found by accessing Services, Run Instrument Import.
- ✓ This utility is now set up to automatically start when AVImark starts, and has the option to Close on AVImark Shutdown.
- ✓ When AVImark opens you will see the “shaking hands” icon located in the System Tray of your workstation. This indicates that AVImark is in communication with the instrument(s) assigned to communication ports on this computer, and is “listening” on each port for any incoming data.
- ✓ The options to Start when AVImark opens, and Close on AVImark Shutdown can be controlled by a right-click option on the icon in the System Tray.




If you have de-selected the Close on AVImark Shutdown option, please note you will need to manually close the connection utility prior to index rebuilding or any other maintenance utility that requires exclusive access to AVImark files. Once the instrument is properly assigned in Hospital Setup, and associated to the AVImark treatment(s), your setup is complete.

Use

When a treatment is entered into Medical History for a patient that has an instrument assignment, AVImark will start listening to the ports in which the associated instruments are assigned for results. As results are ready to be transmitted, AVImark will import the results to the proper Medical History entry. If the proper Medical History cannot be located, the results will be sent to the new “Test Inbox” to await manual import.

First, the import utility looks for a patient with a record number that matches the “Patient ID” field received from the instrument.

The Patient ID that is being searched for a match is the AVImark Patient ID. To keep this ID displayed on the AVImark CID, go into the **Hospital Setup | Advanced Options | Display Patient Record Number** and set it to **True**.



- ✓ When running the test on the designated analyzer, and prompted for a patient identifier, the AVImark Patient ID is the number that will allow the utility to make the proper patient association for automatic result import.
- ✓ If this is successful, then the recent medical history of that patient is searched for a recorded treatment that is set to store results from the instrument.
- ✓ Once the patient identifier is located on the incoming results, the utility will begin searching for a Medical History entry for that patient that is awaiting results.
- ✓ When that Medical History entry is located, the results will be automatically imported into the Tests Attachment area.
- ✓ If the Patient ID doesn't match or exist or there is no matching Medical History entry on that patient awaiting results, the test results are placed in the “Test Inbox.”

- ✓ To manually import the results that were sent to the Test Inbox, simply highlight the Medical History entry you wish to import results for, select the **Tests** attachment | **Import From | Test Inbox**. Here you will be presented with a list of results that were transmitted from instruments that did not find the Medical History entry to complete the automatic import. Select the set of results you wish to import to this history entry, and click Done.
- ✓ When you “Import from” the Test Inbox, the record is actually moved from the Test Inbox onto that Medical History entry. This will allow the Test Inbox to only contain results that have not been imported to Medical History.
- ✓ As an additional safety measure to ensure no results are lost, an RX\$ file is also created with the transmitted results. This RX\$ file is sent to the user defined Instruments storage folder.
- ✓ Results can be imported from this location if needed by opening the Tests attachment, select **Import From | Stored Tests**. This will display all stored RX\$ files on the left pane. You can select the RX\$ file that holds your test, move to the right pane, select the test, **right-click | Import Test**.
- ✓ In the case of an emergency, where access to results is immediately needed, right-click on the treatment in Medical History that has the lab instrument association that represents the test instrument being used, and select Import Results. This will open the Associated Instruments window displaying a list of associated instruments for this treatment. Place a check in the box that represents the instrument from you wish to receive results, click the Import Results button, and AVImark will begin listening to these ports. As the data is sent from the instrument via the port, AVImark will capture and place it immediately in the Test Results window. After the results are received, the Import Results window will disappear, and you can view the results by selecting the Tests attachment option for that Medical History entry.



To alleviate the chance of sending results automatically to the wrong Medical History record that may also meet the requirements for automatic import selection we have incorporated a two day time frame to use for searching, then selecting the oldest entry in that time frame for which these results would apply, to make the result attachment. To summarize, the utility will look for that patient record, then start at the farthest date/time as dictated by the two day time frame, and apply it to the first one it finds. ...versus starting with the most recent and looking back.
